

### CERTIFICATE OF ANALYSIS

PRODUCT NAME. Agar  
 CAS NO. 9002-18-0  
 BATCH NO. 18/MFM/2136  
 MANUFACTURING DATE: May 2018  
 EXPIRY DATE: May 2022

TESTS	SPECS	RESULTS
Appearance	Powder	Conforms
Colour	Cream-white	Conforms
Sieve analysis (ASTM / Retained on Mesh 80)	Max 5%	< 5%
Gel Strength (after autoclave)	>900 - <1050 g/cm2	995
Loss on drying	< 10%	8.03%
pH (1.5% solution) Gel after autoclave	6 – 7.5	6.00
Ash content	<5%	2.6%
Nephelos after autoclaving	< 10 NTU	4.7
Color		
Absorbance 430nm. (in 1.5 sol) After Autoclaving	<0.085	0.047
Melting Point °C	> 85	91.5
Gelling Point °C	34 - 36°C	35
Calcium ppm	<2000	1736
Magnesium ppm	<1000	702
Toxic substances	Absence	Conforms

\* This material is supplied for laboratory use only.

Supplier: Formedium Ltd

Approved by:

Date: 19 JUN 2018  
 For and on behalf of:  
 FORMEDIUM Ltd, Unit 1B  
 Hunstanton Commercial Park  
 PE36 5JQ, England

(Authorised Signature)